

Quality Assurance Certificate

VINOSART PS
Pat.No.US 4,900,449

Order no. 5422506A3
Date of Mfr. 10 / 2010
Pore size 0.45 µm
Lot no. 0145603

This document certifies that the designated product was manufactured by Sartorius in conformance with established Current Good Manufacturing Practice (cGMP) standards.

This product is developed, produced and distributed according to a Quality Management System that is certified for compliance with DIN/ISO 9001.

This product is registered with the Food and Drug Administration (FDA) under the Drug Master File No. 5967.

Complete filter documentation: All data concerning materials used and production data are documented and accessible under the individual batch number which is embossed on every filter element.

Integrity test values: Each membrane filter element has been individually tested for integrity by means of diffusion and each lot randomly by bubble point testing. These tests have been performed according to the procedures stated in the corresponding SOP.

This product is in accordance with the regulation 1935/2004/EC. All raw materials which are used to produce this product are listed in the 2002/72/EC relating to plastic materials intended to come in contact with food.

This product is tested in accordance with the food-and commodities-law of the Federal Republic of Germany (LFGB) section 30 and 31.

For best filtration results, please follow the instructions given in the "Directions for Use".

18.10.2010

Date



Dr. Ralf Wawotzny
Director of Quality Filtration Products



Manufactured by Sartorius Stedim Biotech GmbH
37070 Goettingen, Germany

Phone + 49.551.308.0
Fax + 49.551.308.3289

www.sartorius-stedim.com